



FingerPrint® Sleep Hand-Held Pulse Oximeter

SPECIFICATIONS:

Displays, Indicators & Keys:

SpO₂ LED numeric display
0.300 inch (7.62 mm) high

Pulse Rate LED numeric display
0.300 inch (7.62 mm) high

Pulse Strength Logarithmically scaled, 8-segment bar graph

Keys On/Off and Print

SpO₂:

Range 0-99%

Accuracy ±2% at 70-99%; ≤69% unspecified

Averaging 4 beat average

Pulse Rate:

Range 30 to 254 bpm

Accuracy ±2% or 2 bpm whichever is greater

Averaging 8 second average

Trend Storage:

Parameters SpO₂, Pulse Rate & Relative Time

Interval 4 seconds

Capacity 12 hours of storage

Dimensions:

6.6" H x 2.75" W x 1.43" D
16.7 cm x 7.0 cm x 3.6 cm

Weight:

13 ounces with batteries
369 grams with batteries

Power Requirements:

(4) "AA" Alkaline Batteries

Battery Life:

2 sleep screening cycles (1 cycle = 10 minutes with display on,
8 hours with display off, 5 minutes printing)

or

80 hours Intermittent monitoring (1 min on, 2 min off)

Integrated Printer:

1-inch wide thermal printer paper
15 characters per line

Printer Output:

RS-232C

Mechanical Durability Test Results:

Vibration Testing In compliance with IEC 68-2-34,
IEC 68-2-6 and MIL-STD-810E, Method 514.4

Shock Testing In compliance with IEC 68-2-27

Drop Testing IEC 601-1, Section 21.5

Environmental Specifications:

Operating Temperature 32°F to 131°F (0°C to 55°C)

Storage Temperature -40°F to 167°F (-40°C to 75°C)

Relative Humidity 10-95% (storage)
(non-condensing) 15-95% (operating)

Safety:

IEC 601-1, CSA 601-1, UL 2601, EN865

EMC

IEC601-1-2 CISPR 11 class B

MDD Version Available

ISO 9001 Certified

Catalog Numbers:

3403-000 FingerPrint® Sleep/Summary Printer

Specifications subject to change without notice

Please read the Instructions for Use supplied with the product for detailed instructions, warnings, and cautions.

Q-Line Assistance for U.S.A. Customers

Call our Q-Line at 1-800-348-1842 for emergency and clinical application information 24 hours a day, 365 days a year.
Experienced Clinical Specialists will return your call and assist with any questions you may have.

For more information, or for the name of your BCI representative, contact BCI, Inc.



BCI, Inc.
N7 W22025 Johnson Road Waukesha, Wisconsin 53186-1856 USA
Phone 262-542-3100 Fax 262-542-0718
Toll Free 1-800-558-2345
e-mail: info@smiths-bci.com Internet: http://www.smiths-bci.com

smiths

Smiths Medical – A part of Smiths Group plc